

Certificate of Compliance

Certificate: 80102914 Master Contract: 152450

Project: 80102914 **Date Issued:** 2021-10-27

Issued To: Micro Motion Incorporated

7070 Winchester Cir Boulder, Colorado, 80301

United States

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: Doug Bramwell Doug Bramwell



PRODUCTS

CLASS 2252 06 - PROCESS CONTROL EQUIPMENT CLASS 2252 86 - PROCESS CONTROL EQUIPMENT-Certified to US Standards

Name: 8785 Calibrator

Model(s): 8785 Calibration Standard, 8785 Manufacturing Calibrator

Electrical Rating: 85 Vdc, 2.0 A, Ambient Temperature: +5°C to +40°C



 Certificate: 80102914
 Master Contract: 152450

 Project: 80102914
 Date Issued: 2021-10-27

Conditions of Acceptability:

- 1. The 8785 Calibration Standard and 8785 Manufacturing Calibrator are intended for use with only transmitter models 8782, 8732E, and 8712E.
- 2. The 8785 Calibration Standard and 8785 Manufacturing Calibrator are supplied with and approved for electrical connections with Cable P/N 08785-0507-0001 (Coil Cable) and 08785-0507-0002 (Electrode Cable).

Products may be marked with any of the following Trademarks and/or Tradenames: "Rosemount" or "Micro Motion"

APPLICABLE REQUIREMENTS

CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

ANSI/UL 61010-1 - Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 80102914 Master Contract: 152450

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
80102914	2021-10-27	Models 8785 Calibration Standard and 8785 Manufacturing Calibrator. Transfer from Master Contract 264512, Report 80008851 (Emerson - Rosemount-Micro Motion Inc.) to Master Contract 152450 (Micro Motion Incorporated).