



SIZE	OVERALL END TO END	CENTER TO END	DIAMETER OF FLANGE	CENTER OF HOUSING TO TOP OF SHROUD	CENTER OF HOUSING TO EDGE OF TRANSDUCER CONNECTOR	APPROXIMATE WEIGHT	CLEARANCE FOR SHROUD REMOVAL
	A	B	C	D	E	W.T. (LBS)	G
4"	13.78	6.89	10.00	3.06	4.16	116	6.62
6"	14.00	7.00	12.50	4.22	5.03	163	9.50
8"	16.00	8.00	15.00	5.19	5.88	230	11.25
10"	19.69	9.84	17.50	6.19	6.72	320	13.00

SALES ORDER NO.: _____
 CUSTOMER: _____
 CUSTOMER P.O. NO.: _____
 CERTIFIED CORRECT - MICRO MOTION INC.: _____
 DATE: _____ BY _____

TAG EQUIPMENT NUMBER	SIZE	BORE (F")

Third Angle Projection		CA	02-13-25	PH	CA:1213843	BP	KG
		C	03-12-12	MOD	ECO SP1564	MO	MM
		B	02-28-11	MOD	ECO SP1070	VE	KDG
		A	01-27-11	MOD	RELEASED	MO	KDG
REV	DATE	DWN	DESCRIPTION	CHKD	APPD		

EMERSON

TITLE: OUTLINE & DIMENSIONAL
 3812 LIQUID ULTRASONIC METER
 4" - 10" 300 ANSI RF FLG

DWN MOD DATE 10-25-10 DWG NO. DMC-005633
 CHKD MARK DATE 10-25-10 SHEET 1 OF 1
 APPD KDG DATE 01-27-11

THIS DOCUMENT IS CONFIDENTIAL, CONTAINS PROPRIETARY INFORMATION, AND SHOULD NOT BE DISTRIBUTED, COPIED OR OTHERWISE REPRODUCED WITHOUT THE EXPRESS WRITTEN CONSENT OF MICRO MOTION, INC.